



**CHABOT-LAS POSITAS  
COMMUNITY COLLEGE DISTRICT  
PURCHASING DEPARTMENT**

May 18, 2026

**Addendum No. 1  
INVITATION TO BID.: 25/26-10  
Benchtop Spectrometer**

To: All Prospective Bidders

The following questions were submitted prior to the RFI deadline. Response to the questions are answered herein below:

1. Can you clarify what compatibility with the current user interface means?

***District: Compatibility with the current user interface means that the new spectrometer will have a workflow similar to the workflow of the existing instrument (Nanalysis NMReady 60)***

2. What protocol will be used to determine the sensitivity of the spectrometer? Can you confirm that this will be measured as part of the instrument's acceptance?

***District: Sensitivity and resolution measurements are both provided by the manufacturer. Sensitivity measurement is for 1% ethylbenene, 1 scan.***

3. What protocol will be used to determine the resolution of the spectrometer? Can you confirm that this will be measured as part of the instrument's acceptance?

***District: Resolution measurements are both provided by the manufacturer. We are relying on manufacturer's specifications and will not be performing our own tests.***

---

4. Can you clarify what priority response support means? (found under the warranty description)

***District: Priority response support is remote user support provided in a prompt manner (e.g. same-day or next-day turnaround if possible)***

5. Are there any requirements for the delivery date?

***District: Desired delivery date is before September 2026, but is not a requirement.***

**Marie Hampton,  
Purchasing &  
Warehouse Services  
Manager Chabot-Las  
Positas Community  
College District 7600  
Dublin Blvd. 3<sup>rd</sup> Floor  
Dublin, CA 94568**

***All other terms and conditions remain unchanged.***

Marie Hampton, Purchasing and Warehouse Services  
Chabot-Las Positas Community College District

---

7600 Dublin Blvd., 3rd Floor, Dublin, CA 94568